

CHICAGO STAINLESS EQUIPMENT



1280 S.W. 34th Street • Palm City, Florida • (Phone) 772-781-1441 • (FAX) 772-781-1488 • www.chicagostainless.com

CERTIFICATE OF CONFORMITY (SANI-FLOW BREWERY PRESSURE GAUGE)

Part Number: 2A-E-2U-GF-BT

Serial Number(s): 17032135

We hereby certify that the product contact surface materials in the above referenced parts are AISA Type 316L stainless steel to the specification in the current revision of ASME BPE, section DT, table DT-3.

We also hereby certify that the above referenced gauge as been calibrated by direct comparison to master gauges which are periodically referred to one or more of the primary standards traceable to National Institute of Standards & Technology or other national physical measures recognized by NIST. The above referenced gauge has been calibrated and found to be accurate **+/- 1.5% of full scale** at any point with in its scale.

We also hereby certify that the above referenced gauges are manufactured, assembled and tested in the United States of America at our facility in Palm City, Florida.

Signed:

A handwritten signature in black ink, appearing to read 'Samuel P. Schemenauer', is written over a horizontal line.

3/21/2017

Samuel P. Schemenauer
(Technician)

Date