

**CERTIFICATE OF CONFORMITY
(CHEM-FLOW PRESSURE GAUGE)**

Part Number: 3C-D-15U-GF-BT

Serial Number(s): 13092025

We hereby certify that the base materials in the above referenced parts are AISA Type 316L stainless steel to the specification in the current revision of ASME BPE, section DT, table DT-3. The metallic black coating bonded to the base material is DuPont Teflon® 420-109 which complies with FDA regulations in CFR 21 governing components of coatings for direct food contact. The coating has a maximum use temperature of 500°F (260°C) and thickness of approximately 0.5 to 1 mils (500 to 1,000µin or 13 to 25µm).

We also hereby certify that the above referenced gauge has been calibrated by direct comparison to master gauges which are periodically referred to one or more of the primary standards traceable to National Institute of Standards & Technology or other national physical measures recognized by NIST. The above referenced gauge has been calibrated and found to be accurate **+/- 3/4% of full scale** at any point within its scale. The accuracy has been certified at every major subdivision on the dial face.

We also hereby certify that the above referenced gauges are manufactured, assembled and test in the United States of America at our facility in Palm City, Florida.

Signed:



**Samuel P. Schemenauer
(Technician)**

09/20/2013

Date