

**CERTIFICATE OF CONFORMITY
(SANI-FLOW DIGITAL THERMOMETER)**

Part Number: DT-2U-BT-DF

Serial Number(s): 23022108

We hereby certify that the product contact surface materials in the above referenced parts are AISA Type 316L stainless steel to the specification in the current revision of ASME BPE, section DT, table DT-3. That the finished part has been mechanically polished to a minimum $R_{a\max}$ surface finish of 8 microinches.

We also hereby certify that the above referenced thermometer has been calibrated by direct comparison to master gauges which are periodically referred to one or more of the primary standards traceable to National Institute of Standards & Technology or other national physical measures recognized by NIST. The above referenced gauge has been calibrated and found to be accurate **+/- 0.5°F (0.3°C)** at any point within its scale.

We also hereby certify that the above referenced gauges are manufactured, assembled and test in the United States of America at our facility in Palm City, Florida.

Signed:



**Samuel P. Schemenauer
(Technician)**

February 21, 2023

Date