



Certificate of Calibration

Report number: 44972 -2

Test Result: **PASS**

Instrument Under Test (IUT)			
Manufacturer:	Chicago Stainless Equipment, Inc.		
Description:	Sani-Flow Digital Thermometer	Calibration Date:	2/15/2023
Model Number:	DT-1U-BT-DF	Expiration Date:	2/15/2024
Serial Number:	23021420	ESN:	11784

This certification was made by direct comparison to master gauges which are periodically referred to one or more of the primary standards traceable to National Institute of Standards & Technology or other national physical measures recognized by NIST. Reference Standards used in this calibration are traceable to the National Institute of Standards and Technology of the United States, through the following report numbers:

Reference Standards			
Description:	IRTD-400	Reference Thermometer	Reference Thermometer
Manufacturer:	Kaye - GE	Fluke	Fluke
Model Number:	M2801	1551-A	1551-A
Serial Number:	50575	4599010	4527008
Report Number:	22037900	4465-1	4465-2
Due Date:	11/15/2023	4/14/2023	4/14/2023
Accuracy:	±0.04°C	±0.05°C	±0.05°C

Test Results

Laboratory ambient conditions throughout this calibration were: 73 to 80°F (23 to 27°C)

Stated Accuracy of Instrument Under Test (IUT): ±0.50°F				
Reference Reading	IUT Reading	Difference	Allowable Error	Condition
±32.00°F	±32.00°F	±0.00°F	±0.50°F	PASS
±125.00°F	±125.00°F	±0.00°F	±0.50°F	PASS
±212.00°F	±212.00°F	±0.00°F	±0.50°F	PASS
±300.00°F	±299.80°F	±0.20°F	±0.50°F	PASS


 Sam Schemenauer
 Calibration Technician


 Greg Mizarz
 Quality Representative